

计量标准考核证书

Certificate for Examination of Measurement Standard

(2022) 椒 量标 企 证字第 003 号

根据《中华人民共和国计量法》，按照《计量标准考核规范》的要求，考核合格。特发此证。

This is to certify that the measurement standard conforms with requirements of the "Rule for the Examination of Measurement Standards" according to "the Law on Metrology of the People's Republic of China".

建标单位 浙江鑫泰检测技术有限公司

Possessor of the Measurement Standard

计量标准名称 环境试验设备温度、湿度校准装置 代码 04117300

Name of Measurement Standard Code

测量范围 温度：(-60~300)℃；
湿度：(40~90)%RH

Measuring Range

不确定度或准确度等级

或最大允许误差

温度：U=0.10℃，k=2；
湿度：U=1.1%RH，k=2

Uncertainty / Accuracy Class /

Maximum Permissible Error

保存地点：浙江省台州市椒江区下陈街道聚星科创园 60 幢 1 号二楼热工实验室
Installed in

发证机关（印章）

The Issuing Authority

发证日期 2022 年 01 月 20 日

Date Issued

有效期至 2027 年 01 月 19 日

Date of Expiry

计量标准器 Measurement Standard	名称 Name	型号 Type	测量范围 Measuring Range	不确定度或准确度等级 或最大允许误差 Uncertainty or Accuracy Class or Maximum Permissible Error	制造厂及 出厂编号 Manufacturer and Series Number
	温湿度巡检仪	DTZ-300	温度： (-60~300)℃	温度： U=0.10℃，k=2	泰安德图自动化 仪器有限公司/ DW-2067
			湿度： (40~90)%RH	湿度：U=1.1%RH，k=2	
主要 配套 设备 Main Auxiliary Equipment	/	/	/	/	/
开展的 检定或 校准项目 Verification or Calibration Items	名 称 Name	测量范围 Measuring Range	不确定度或准确度等级 或最大允许误差 Uncertainty or Accuracy Class or Maximum Permissible Error		依据的计量检定规程或 计量技术规范的编号及 名称 Verification Regulation or Measurement Technical Specification and Its Code
	温度环境试验 设备	温度范围： (-60~300)℃	(-60~200)℃： 温度偏差：±2.0℃，温度均匀度：2.0℃， 温度波动度：±0.5℃； (200~300)℃： 温度偏差：±3.0℃，温度均匀度：3.0℃， 温度波动度：±1.0℃及以下		JJF 1101-2019 《环境试验设备温 度、湿度参数校准规 范》
	湿热环境试验 设备	温度范围： (10~85)℃ 湿度范围： (40~90)%RH	(10~85)℃、(40~75)%RH： 温度偏差：±2.0℃，温度均匀度：2.0℃， 温度波动度：±1.0℃；湿度偏差：±5.0%RH， 湿度均匀度：7.0%RH，湿度波动度： ±3.0%RH (10~85)℃、(75~90)%RH： 温度偏差：±2.0℃，温度均匀度：2.0℃， 温度波动度：±0.5℃；湿度偏差：±3.0%RH， 湿度均匀度：5.0%RH，湿度波动度： ±3.0%RH 及以下		